



Labor Dr.Hüfner GmbH

Specification – reference values
Cell Count Standard - MS 04 – 2024

Cow Milk – Cell Count Standards			
	Milk I “low”	Milk II “high”	Milk III “Extra-high”
Batch	04-2024	04-2024	04-2024
somatic cells x 1000	152	380	650
uncertainty of measurement -measuring range	136-167	342-418	585-715

Goat Milk – Special Cell Count Standards			
	Goat Milk “low”	Goat Milk “high”	Goat Milk “Extra-high”
Batch	March 24	March 24	Febr. 24
somatic cells x 1000	1050	1.510	2.020
uncertainty of measurement -measuring range	893-1208	1284-1737	1717-2323

Hergatz, 11.04.2024

Head of the Institute

Durch die DAkKS nach DIN EN ISO 17034
akkreditierter Referenzmaterialhersteller



Erstellt von/am:	Hü / Mai 2014	Geändert von/am:	Bo / 13.06.23	Freigabe von/am:	Bo / 19.06.23
Geprüft von/am:	Hü / 11.09.19	Geprüft von /am	Hü / 19.06.23		
		Specification Cell Count Standard		Seite: 1 von 1 Version: 04	