



Specification – reference values
Milk Standard - MS 04 – 2024

(A) BactoScan FC				
	MS A	MS B	MS C*)	RELATION MS B/MS A
Batch	MS 04	MS 04	MS 04	
Reference value MS 04 Impulses	180	530	1000	2,95
uncertainty of measurement -measuring range	162-198	477-583	900-1100	

(B) Bactocount (BC)				
	MS A	MS B	MS C*)	RELATION MS B/MS A
Batch	MS 04	MS 04	MS 04	
Reference value MS 04 Impulses	221	652	1230	2,95
uncertainty of measurement -measuring range	188-254	554-750	1107-1353	

Hergatz, den 11.04.2024

Dr. Josef Hüfner

Durch die DAkkS nach DIN EN ISO 17034
akkreditierter Referenzmaterialhersteller



*) currently not part of the accreditation certificate

Erstellt von/am:	Hü / Mai 2014	Geändert von/am:	Bo / 16.06.23	Freigabe von/am:	Bo / 19.06.23
Geprüft von/am:	Hü / 11.09.19	Geprüft von /am	Hü / 19.06.23		
MIH		Specification BSC-FC,BC Milk Standard			Seite: 1 von 1 Version: 04